

The following errors exist in the published version of “Model-based iterative reconstruction for flat-panel cone-beam CT with focal spot blur, detector blur, and correlated noise,” DOI 10.1088/0031-9155/61/1/296

1. The fidelity terms in equations 10 and 12 are missing a multiplication by 0.5
2. Equation 14 should be

$$\mu(x_j) = a + b \operatorname{erf} \left( \frac{2\sqrt{\log(2)}(x_j - d)}{\text{FWHM}} \right)$$

3. In section 3.2 a reference to Figure 10(e) should be 9(f)